

Summary Table

3/12/07 12:04 PM

Table #

Title

Characteristic	Statistic	Nominal Concentration (ng/mL)				
		QC1 2.0	QC2 10.0	QC3 32.0	QC4 50.0	QC5 64.0
# Results	N	18	18	18	17	18
Accuracy	Mean Bias (%RE)	-2.9	-7.4	-1.1	-5.0	-2.0
	LCL	-12.157	-15.608	-7.634	-14.592	-11.800
	UCL	6.379	0.863	5.342	4.692	7.807
Precision	Intrabatch (%CV)	6.9	7.5	9.6	6.1	5.5
	Interbatch (%CV)	10.5	10.0	10.0	10.4	10.4
Accuracy + Precision	 Mean + Interbatch	13.363	17.326	11.133	15.356	12.353
90% Expectation	Lower Limit (%RE)	-23.2	-26.2	-19.1	-25.5	-22.9
Tolerance Interval	Upper Limit (%RE)	17.4	11.4	16.8	15.6	18.9